

CE DECLARATION OF CONFORMITY

NOT TRANSFERABLE

Responsible Company: Distek, Inc.
121 North Center Dr.
North Brunswick, NJ. 08902

Equipment Type & Model: Distek Autosamplers:
DS 4300
Evolution 4300
Syringe Pump

Equipment Classification: Electrical equipment for measurement,
control and laboratory use.

Declared Compliance & Test Standards: EMC Directive 89/336/EEC/93/68/EEC
IEC 61326:2002
EN 61000-3-2
EN 61000-3-3
EN 61010-1:2001 (2nd Edition)
IEC 1010-2-010:92 (as applicable)
UL 61010A-1
CAN/CSA 1010.1-92



Jeff Brinker
Engineering Manager
November 11, 2003



121 North Center Drive
North Brunswick, New Jersey 08902