

CE DECLARATION OF CONFORMITY

NOT TRANSFERABLE

Responsible Company: Distek, Inc.
121 North Center Dr.
North Brunswick, NJ. 08902

Equipment Type & Model: Distek Opt-Diss 410, Fiber Optic
Dissolution Measuring Instrument

Equipment Classification: Electrical equipment for measurement,
control and laboratory use

Declared Compliance & Test Standards: EMC Directive 2014/30/EU
EN 61326-1 (2013)
FCC CFR 47, Part 15, Subpart B, 2012
Low Voltage Directive 2014/35/EU
UL 61010-1:2012 Ed.3 & R:15 July 2015
CSA C22.2 #61010-1: 2012 3rd Edition
IEC 61010-1: 2017, Edition 3.1



Peter K. Schutz
Engineering Manager
February 8, 2017



121 North Center Drive
North Brunswick, New Jersey 08902